



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

Skeletal Dynamics, LLC  
Ms. Ana Escagedo  
President  
8905 SW 87 Avenue, Suite 201  
Miami, Florida 33176

June 27, 2017

Re: K171590  
Trade/Device Name: Distal Elbow Plating System  
Regulation Number: 21 CFR 888.3030  
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And  
Accessories  
Regulatory Class: Class II  
Product Code: HRS, HWC  
Dated: May 30, 2017  
Received: May 31, 2017

Dear Ms. Ana Escagedo:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR

Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K171590

Device Name

Distal Elbow Plating System

Indications for Use (Describe)

The Skeletal Dynamics Distal Elbow Plating System is intended for fixation of fractures, fusions, osteotomies and non-unions of the radius and ulna, particularly in osteopenic bone.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*



**510(k) Summary of Safety and Effectiveness  
Skeletal Dynamics Distal Elbow Plating System**

June 22, 2017

**Submitter:**

Skeletal Dynamics, LLC  
8905 SW 87 Avenue  
Suite 201  
Miami, FL 33176  
Tel: (305) 596-7585  
Fax: (305) 596-7591  
Contact: Ana M. Escagedo, Vice President  
Email: [aescagedo@skeletaldynamics.com](mailto:aescagedo@skeletaldynamics.com)

Establishment Registration Number: 3006742481

**Name and Classification:**

Name	Distal Elbow Plating System
Common Name	Plate, fixation, bone
Classification	21 CFR §888.3030
Product Code	HRS, HWC
Class	Class II

**Predicate Devices:**

Skeletal Dynamics Distal Elbow Plating System  
(K140892) DePuy Anatomic Plating System (K082300)

**Description of the Device:**

The Skeletal Dynamics Distal Elbow Plating System is comprised of:

- Titanium alloy plates and screws in various configurations
- Cobalt chrome polyaxial screws
- Stainless steel K-wires (for provisional fixation not for implantation)
- System specific instrumentation

The current Skeletal Dynamics Distal Elbow Plating System (K140892) includes the radial head and proximal ulna plates. Plates in double hockey stick, Y and coronoid shape configurations with the same node and inter-node features as the radial head plate are being added for use in the proximal radius and ulna in this submission. All plates are made of medical grade titanium alloy. The radial head, proximal ulna and coronoid plates are offered in left and right applications, the proximal ulna plate is provided in various lengths, and the double hockey stick and Y plates are in single size and have bilateral application. Additional screws length as well as new threaded pegs, washers, polyaxial locking screws, and instruments

have been added in the subject system in this submission. The system is provided non-sterile and is sterilized in the user facility.

**Intended Use:**

The Skeletal Dynamics Distal Elbow Plating System is intended for fixation of fractures, fusions, osteotomies and non-unions of the radius and ulna, particularly in osteopenic bone.

**Summary of Technological Characteristics / Substantial Equivalence:**

The substantial equivalence of the Skeletal Dynamics Distal Elbow Plating System to the predicate device is demonstrated by similarities in intended use, indications for use, materials, design (fundamental scientific technology), performance, sterility and packaging, and does not present any new issues of safety or effectiveness.

**Performance Testing:**

Engineering analysis and preclinical testing demonstrated that the Skeletal Dynamics Distal Elbow Plating System is substantially equivalent to the predicate device currently marketed. Mechanical testing which established equivalency included conformance to ASTM F382-14, Standard Specification and Test Methods for Metallic Bone Plates and ASTM F543-17, Standard Specification and Test Methods for Metallic Bone Screws.

**Conclusion:**

The Skeletal Dynamics Distal Elbow Plating System is substantially equivalent to the predicate devices identified in this premarket notification.