

REDUCT[®] Headless Compression Screw System

INSTRUCTIONS FOR USE

R: For use by physicians only. Federal Law restricts this device to sale by or on the order of a physician.

Failure to follow instructions may lead to patient injury.

This package insert is designed to provide Instructions for Use of the REDUCT[®] Headless Compression Screw System; it is not a reference to surgical techniques.

Description

The REDUCT[®] Headless Compression Screw (HCS) System consists of the following screws from medical grade titanium alloy (ASTM F-136).

- 2.5mm cannulated HCS screws: 10mm - 30mm
- 3.5mm cannulated HCS screws: 10mm - 50mm
- 4.5mm cannulated HCS screws: 20mm - 65mm
- 2.0mm non-cannulated HCS Arthrodesis screws: 20mm - 44mm
- 2.5mm cannulated HCS Arthrodesis screws: 26mm - 40mm
- 3.5mm cannulated HCS Arthrodesis screws: 32mm - 46mm

The REDUCT[®] Headless Compression System includes instrumentation identified for the associated surgical techniques. Both the REDUCT[®] Headless Compression Screws and instrumentation are provided non-sterile and must be sterilized in the user facility.

Indications

The Skeletal Dynamics REDUCT[®] Headless Compression Screw System is intended for fixation of osseous fragments or fractures, arthrodesis of small joints, and osteotomies, with the appropriately sized screw.

Contraindications

Prior to using the REDUCT[®] Headless Compression Screw System, ensure that none of the following patient conditions are present: active or latent infection, sepsis, osteoporosis, insufficient quantity or quality of bone and/or soft tissue, material sensitivity (if sensitivity is suspected, tests are performed prior to implantation), or patients who are unwilling or incapable of following post operative care instructions. These devices are not intended for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic, or lumbar spine.

Warnings

- The patient must be cautioned, preferably in writing, about the use, limitations, and potential adverse effects of this device including the possibility of delayed union, non-union, device or treatment failure as a result of loose fixation and/or loosening, stress, excessive activity, or weight bearing or load bearing, and the possibility of nerve or soft tissue damage related to either surgical trauma or the presence of the device.
- The patient should be informed about the importance of following the prescribed post-operative rehabilitation protocol and to understand the possible limitations in activities of daily living. The patient must be warned that failure to follow postoperative care instructions may cause the implant or treatment to fail.
- For safe effective use of the implant, the surgeon must be thoroughly familiar with the surgical technique for the device, implant, and associated instruments. Potential failures of the Headless Compression Screw System may include delayed union, non-union, loosening of fixation, stress fractures of the bones, or incomplete healing as a result of excessive activity, overloading or non compliance to post operative rehabilitation.
- The device is not designed to withstand the stress of weight bearing, load bearing, or excessive physical activity. Device breakage may occur when the implant is subjected to excessive loading associated with delayed union or nonunion. Improper insertion of the device during implantation may also increase the possibility of loosening, or migration.
- DO NOT reuse any of the REDUCT[®] Headless Compression Screw System's implantable components. Reuse may compromise the structural integrity of the screw and/or lead to failure, which may result in patient injury.
- Seek medical help immediately if implant malfunctions.

- The system is to be used only with Skeletal Dynamics instruments, implants and accessories.

Precautions

- Protect the System's implantable components against scratching or nicking. Such stress concentration can lead to implant failure.
- Before using the REDUCT® Headless Compression Screw System, inspect all implants and instruments for wear, disfiguration and physical damage. If evidence of wear, disfiguration or physical damage is found, DO NOT use and contact your local Skeletal Dynamics representative or the Skeletal Dynamics Customer Care Department.
- DO NOT permanently implant the Skeletal Dynamics K-Wires; they are only intended to be used for provisional fixation and guidance.
- K-Wires are double trocar. User should handle K-Wires accordingly during insertion and removal to prevent unintended K-Wire penetration or injury.
- Do not mix implant components from different manufacturers for metallurgical, biomechanical and functional reasons.
- DO NOT use screw lengths that will excessively protrude through the far cortex as it may result in soft tissue irritation.
- The benefits from implant surgery may not meet the patient's expectations or may deteriorate over time, requiring revision surgery to replace the implant or to carry out alternative procedures. **Note:** To maintain traceability of the implantable components, record each of the respective components LOT numbers in the patient records post implantation.

MRI Safety Information

The REDUCT® Headless Compression Screw System has not been evaluated for safety and compatibility in the MR environment. It has not been tested for heating, migration, or image artifact in the MR environment. The safety of REDUCT® Headless Compression Screw System in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

Potential Adverse Events

Possible adverse effects associated with headless compression screws are infection, pain, stiffness, discomfort, or abnormal sensations and nerve or soft tissue damage due to the use of an implant or due to surgical trauma. The implant may break due to excessive activity, prolonged loading, incomplete healing, or excessive force on the implant during insertion. Metal sensitivity or histological or allergic or adverse foreign body reaction resulting from implantation of a foreign material may occur. Nerve or soft tissue damage, necrosis of the tissue or inadequate healing may result from the presence of an implant or due to surgical trauma.

Directions for Use

The REDUCT® Headless Compression Screw System should only be used by surgeons who have experience with this system. Each surgeon must evaluate the appropriateness for the use of the Headless Compression Screw System based on their clinical experiences.

The surgeon should select the type and size implant to best meet the patient's needs. Although the surgeon is the medical intermediary between the company and the patient, this document contains important medical information provided in this document should be shared to the patient.

It is the responsibility of the surgeon to be familiar with the procedure before use of this device. Additionally, it is the responsibility of the surgeon to be familiar with relevant publications regarding the procedure prior to use. Please refer to the Headless Compression Screw System Surgical Technique Guide to review the surgical approach as described by Jorge L. Orbay, M.D. of the *Miami Hand Institute* located in Miami, Florida.

Cleaning

Upon receipt by the user facility, the REDUCT® Headless Compression Screw System should be cleaned prior to sterilization. The recommended manual cleaning instructions are set forth below. Other cleaning methods must be validated by the user.

Implant Cleaning

Implanted plates, screws, or associated components should never be re-used. After each use, unused implants must be cleaned separately from contaminated instruments to prevent cross-contamination utilizing the cleaning instructions provided below.

Warnings & Precautions

- If the implant has been in contact with the patient, body fluids or tissues or is damaged, it may NOT be reprocessed and MUST be properly discarded.
- Users should wear appropriate personal protective equipment (PPE).
- Users should be qualified personnel with documented evidence of training and competency. Training should be inclusive of current applicable guidelines and standards and healthcare facility policies.

Instrument Cleaning

The REDUCT® Headless Compression Screw System instrumentation must be cleaned thoroughly before re-use to achieve sterilization.

Warnings & Precautions

- The System's reusable instruments and accessories, including sterilization tray and tray components, should be decontaminated immediately after completion of the surgical procedure. Contaminated instruments should not be allowed to dry prior to cleaning/reprocessing. Excess blood or debris should be wiped off to prevent it from drying.
- Only qualified personnel with documented evidence of training and competency should clean the instruments. Training should be inclusive of current applicable guidelines and standards and healthcare facility policies.
- Avoid the use of metal brushes or scouring pads during the cleaning process.
- Instruments should be rinsed of cleaning agents to prevent residue.
- Do not use mineral oil or silicone lubricants on instruments.
- Neutral pH enzymatic and cleaning agents are recommended for cleaning instruments. It is important that alkaline cleaning agents are thoroughly neutralized and rinsed from instruments.
- Prior to sterilization, instruments should be inspected for cleanliness of surfaces, joints, and lumens, proper function, and wear and tear. If the product cannot be cleaned after repeated washing or If evidence of wear, disfiguration or physical damage is found, DO NOT use and contact your local Skeletal Dynamics representative or the Skeletal Dynamics Customer Care Department.

Cleaning Instructions

Cleaning should begin at the point of use prior to processing. Keep instruments moist after use to prevent soil from drying on them. An enzymatic detergent (Enzol) was used to validate the cleaning process.

1. Disassemble instrumentation, if applicable.
2. Rinse all components, including instruments, sterilization tray and tray components, thoroughly under running cool tap water. While rinsing, use a soft bristle brush to loosen and remove as much visible soil as possible from components.
3. Soak all components in a neutral enzymatic cleaner for a minimum of ten (10) minutes. Components must be fully immersed in the cleaner. Follow the cleaner manufacturer's instructions for cleaner preparation and whenever longer exposure times are recommended.
4. Thoroughly rinse all components with cool water. While rinsing, use soft bristle brushes, pipettes or a water jet to clean out lumens, holes, and other challenging features.
5. Manually scrub all components thoroughly in newly made, clean, neutral pH enzymatic cleaner using soft bristle brushes or pipettes. All lumens, holes, hinged components, mating surfaces, and crevices, and challenging components should be thoroughly scrubbed. Actuate all moveable features and expose all areas to cleaner and to the brush or pipette.
6. Rinse all components thoroughly under reverse osmosis/deionized (RO/DI) water; using pipettes or a water jet to clean out lumens, holes, and other hard to reach or challenging features. Actuate all movable features to fully irrigate all areas.
7. Visually inspect all components for soil. Repeat the cleaning procedure until no visible soil remains on the components.
8. Perform a final rinse on all components using running RO/DI water.

9. Dry the clean components using compressed air or a soft, lint free, clean cloth.

Sterilization

The REDUCT® Headless Compression Screw System is provided non-sterile. This system is intended for steam sterilization at the healthcare facility.

1. Place all components and accessories into the designated areas of the sterilization tray. Do not stack trays during sterilization.
2. Steam sterilization may be accomplished using one of the cycles shown below:

Cycle Times for Dynamic-Air-Removal (Vacuum) Steam Sterilization Cycles

Item	Exposure time at 132°C (270° F)	Drying Time
Wrapped Sterilization Tray	4 minutes (wrapped)	40 minutes

- Follow ANSI/AAMI ST79:2006, Comprehensive guide to steam sterilization and sterility assurance in health care facilities.
- Immediate-Use Steam Sterilization (IUSS) not recommended.
- Usage of an FDA cleared wrap is required.
- Subsequent instrument sterilization needs to be performed in the tray system provided. For reuse and sterilization, instruments should be arranged within the tray system in the manner supplied by the company.

Storage

When not in use, store the clean and disinfected REDUCT® Headless Compression Screw System within the Sterilization Tray. Store in a cool dry place and keep away from direct sunlight. Prior to use, inspect the instrumentation for serviceability.

Disclaimer of Warranty and Limited Remedies

Skeletal Dynamics, Inc. makes no express or implied warranty, including any implied warranty of merchantability or fitness for a particular purpose, on the product(s) described in this publication. Skeletal Dynamics, Inc. shall not be liable under any circumstances for any direct, incidental or consequential damages other than as expressly provided by specific law. No person has authority to bind Skeletal Dynamics, Inc. to any representation or warranty except as specifically set forth in this publication. Descriptions or specifications provided by Skeletal Dynamics, Inc. in any publication are only included to generally describe the product when manufactured and do not constitute any express warranties.



Skeletal Dynamics Inc
7300 N. Kendall Dr. / Suite 800
Miami, FL 33156
1-877-753-5396



Emergo Europe.
Westervoortsedijk 60
6827 AT Arnhem
The Netherlands



REDUCT® Headless Compression Screw System

Inventory Control Sheet

2.5mm Headless Compression Screw	
Screw, Headless Compression, 2.5mm x 10mm, Ti HCS-25010 (01)00841506101835  (01) 00841506101835	Screw, Headless Compression, 2.5mm x 18mm, Ti HCS-25018 (01)00841506101873  (01) 00841506101873
Screw, Headless Compression, 2.5mm x 11mm, Ti HCS-25011 (01)00841506115207  (01) 00841506115207	Screw, Headless Compression, 2.5mm x 20mm, Ti HCS-25020 (01)00841506101880  (01) 00841506101880
Screw, Headless Compression, 2.5mm x 12mm, Ti HCS-25012 (01)00841506101842  (01) 00841506101842	Screw, Headless Compression, 2.5mm x 22mm, Ti HCS-25022 (01)00841506101897  (01) 00841506101897
Screw, Headless Compression, 2.5mm x 13mm, Ti HCS-25013 (01)00841506115214  (01) 00841506115214	Screw, Headless Compression, 2.5mm x 24mm, Ti HCS-25024 (01)00841506101903  (01) 00841506101903
Screw, Headless Compression, 2.5mm x 14mm, Ti HCS-25014 (01)00841506101859  (01) 00841506101859	Screw, Headless Compression, 2.5mm x 26mm, Ti HCS-25026 (01)00841506101910  (01) 00841506101910
Screw, Headless Compression, 2.5mm x 15mm, Ti HCS-25015 (01)00841506115221  (01) 00841506115221	Screw, Headless Compression, 2.5mm x 28mm, Ti HCS-25028 (01)00841506101927  (01) 00841506101927
Screw, Headless Compression, 2.5mm x 16mm, Ti HCS-25016 (01)00841506101866  (01) 00841506101866	Screw, Headless Compression, 2.5mm x 30mm, Ti HCS-25030 (01)00841506101934  (01) 00841506101934
Screw, Headless Compression, 2.5mm x 17mm, Ti HCS-25017 (01)00841506115238  (01) 00841506115238	
3.5mm Headless Compression Screw	
Screw, Headless Compression, 3.5mm x 10mm, Ti HCS-35010 (01)00841506101941  (01) 00841506101941	Screw, Headless Compression, 3.5mm x 20mm, Ti HCS-35020 (01)00841506101996  (01) 00841506101996
Screw, Headless Compression, 3.5mm x 11mm, Ti HCS-35011 (01)00841506115245  (01) 00841506115238	Screw, Headless Compression, 3.5mm x 22mm, Ti HCS-35022 (01)00841506102009  (01) 00841506102009

REDUCT® Headless Compression Screw System

Inventory Control Sheet

Screw, Headless Compression, 3.5mm x 12mm, Ti HCS-35012 (01)00841506101958  (01) 00841506101958	Screw, Headless Compression, 3.5mm x 24mm, Ti HCS-35024 (01)00841506102016  (01) 00841506102016
Screw, Headless Compression, 3.5mm x 13mm, Ti HCS-35013 (01)00841506115252  (01) 00841506115252	Screw, Headless Compression, 3.5mm x 26mm, Ti HCS-35026 (01)00841506102023  (01) 00841506102023
Screw, Headless Compression, 3.5mm x 14mm, Ti HCS-35014 (01)00841506101965  (01) 00841506101965	Screw, Headless Compression, 3.5mm x 28mm, Ti HCS-35028 (01)00841506102030  (01) 00841506102030
Screw, Headless Compression, 3.5mm x 15mm, Ti HCS-35015 (01)00841506115269  (01) 00841506115269	Screw, Headless Compression, 3.5mm x 30mm, Ti HCS-35030 (01)00841506102047  (01) 00841506102047
Screw, Headless Compression, 3.5mm x 16mm, Ti HCS-35016 (01)00841506101972  (01) 00841506101972	Screw, Headless Compression, 3.5mm x 35mm, Ti HCS-35035 (01)00841506105840  (01) 00841506105840
Screw, Headless Compression, 3.5mm x 17mm, Ti HCS-35017 (01)00841506115276  (01) 00841506115276	Screw, Headless Compression, 3.5mm x 40mm, Ti HCS-35040 (01)00841506105857  (01) 00841506105857
Screw, Headless Compression, 3.5mm x 18mm, Ti HCS-35018 (01)00841506101989  (01) 00841506101989	Screw, Headless Compression, 3.5mm x 45mm, Ti HCS-35045 (01)00841506105864  (01) 00841506105864
4.5mm Headless Compression Screw	
Screw, Headless Compression, 4.5mm x 20mm, Ti HCS-45020 (01)00841506115771  (01) 00841506115771	Screw, Headless Compression, 4.5mm x 30mm, Ti HCS-45030 (01)00841506108711  (01) 00841506108711
Screw, Headless Compression, 4.5mm x 22mm, Ti HCS-45022 (01)00841506115788  (01) 00841506115788	Screw, Headless Compression, 4.5mm x 35mm, Ti HCS-45035 (01)00841506108728  (01) 00841506108728
Screw, Headless Compression, 4.5mm x 24mm, Ti HCS-45024 (01)00841506108681  (01) 00841506108681	Screw, Headless Compression, 4.5mm x 40mm, Ti HCS-45040 (01)00841506108735  (01) 00841506108735
Screw, Headless Compression, 4.5mm x 26mm, Ti HCS-45026 (01)00841506108698  (01) 00841506108698	Screw, Headless Compression, 4.5mm x 45mm, Ti HCS-45045 (01)00841506108742  (01) 00841506108742
















REDUCT® Headless Compression Screw System

Inventory Control Sheet

Screw, Headless Compression, 4.5mm x 28mm, Ti HCS-45028 (01)00841506108704  (01) 00841506108704	Screw, Headless Compression, 4.5mm x 50mm, Ti HCS-45050 (01)00841506108759  (01) 00841506108759
Single Use (Disposable) Instruments	
Drill, Quick Connect, 1.9mm Cannulated DRLL-CDC-19 (01)00841506101200  (01) 00841506101200	HCS Driver, 2.5mm DRVR-HCS-0915 (01)00841506101309  (01) 00841506101309
Drill, Quick Connect, 2.7mm Cannulated DRLL-CDC-27 (01)00841506101217  (01) 00841506101217	HCS Driver, 3.5mm DRVR-HCS-1420 (01)00841506101316  (01) 00841506101316
Drill, Quick Disconnect, 3.7 Cannulated DRLL-CDC-37 (01)00841506109657  (01) 00841506109657	HCS Driver, 4.5mm DRVR-HCS-1425 (01)00841506108766  (01) 00841506108766
Drill, Countersink, 2.7mm, Cannulated DRLL-CSK-27 (01)00841506103846  (01) 00841506103846	K-Wire, Double Trocar, 0.9mm x 152mm KWIR-DT-09152 (01)00841506115320  (01) 00841506115320
Drill, Countersink, 3.5mm Long, Cannulated DRLL-CSK-35L (01)00841506116211  (01) 00841506116211	K-Wire, Double Trocar, 1.4mm x 165mm KWIR-DT-14165 (01)00841506115337  (01) 00841506115337
Drill, Countersink, 4.5, Cannulated DRLL-CSK-45 (01)00841506108780  (01) 00841506108780	K-Wire, Single Trocar, 0.9mm x 152mm KWIR-ST-09152 (01)00841506115283  (01) 00841506115283
K-Wire, 0.9mm x 152mm, Double Trocar KWIR-HCS-09152 (01)00841506109145  (01) 00841506109145	K-Wire, Single Trocar, 1.4mm x 165mm KWIR-ST-14165 (01)00841506115290  (01) 00841506115290
K-Wire, 1.4mm x 165mm, Double Trocar KWIR-HCS-14165 (01)00841506102481  (01) 00841506102481	Drill, Countersink, 3.5mm, Cannulated DRLL-CSK-35 (01)00841506103853  (01) 00841506103853

REDUCT® Headless Compression Screw System

Inventory Control Sheet

Reusable Instruments			
Depth Gauge, Universal, 50mm DPGA-UNV-050 (01)00841506106656	 (01) 00841506106656	Handle, Small QC, Fixed HNDL-SQC-FXD (01)00841506102078	 (01) 00841506102078
Large Reduction Forceps FRCP-BHL-RTP (01)00841506107264	 (01) 00841506107264	HCS Depth Gage, 2.5mm VHCS-DGA-25 (01)00841506103716	 (01) 00841506103716
HCS Tissue Protector HCS-TPG (01)00841506105437	 (01) 00841506105437	HCS Depth Gage, 3.5mm, Long VHCS-DGA-35L (01)00841506117744	 (01) 00841506117744
HCS Obturator HCS-TPO (01)00841506105420	 (01) 00841506105420	HCS Depth Gage, 4.5 VHCS-DGA-45 (01)00841506108797	 (01) 00841506108797
HCS Wire Pusher HCS-WP (01)00841506105413	 (01) 00841506105413	HCS Depth Gage, 3.5mm VHCS-DGA-35 (01)00841506103723	 (01) 00841506103723
Depth Gauge, Universal, 30mm DPGA-UNV-030 (01)00841506101194	 (01) 00841506101194	HCS K-Wire Needle HCS-KWN (01)00841506105413	 (01) 00841506105413
HCS Parallel K-Wire Guide, 1 HCS-PWG-01 (01)00841506131061	 (01) 00841506131061	HCS Parallel K-Wire Guide, 2 HCS-PWG-02 (01)00841506131078	 (01) 00841506131078
HCS Parallel K-Wire Guide, 3 HCS-PWG-03 (01)00841506131085	 (01) 00841506131085		

REDUCT® Arthrodesis Screw System

Inventory Control Sheet

2.0mm Arthrodesis Screw			
Screw, Arthrodesis, 2.0 x 20mm, Ti HCSD-20020 (01)00841506112749	 (01) 00841506112749	Screw, Arthrodesis, 2.0 x 28mm, Ti HCSD-20028 (01)00841506115528	 (01) 00841506115528
Screw, Arthrodesis, 2.0 x 22mm, Ti HCSD-20022 (01)00841506115498	 (01) 00841506115498	Screw, Arthrodesis, 2.0 x 30mm, Ti HCSD-20030 (01)00841506108902	 (01) 00841506108902
Screw, Arthrodesis, 2.0 x 24mm, Ti HCSD-20024 (01)00841506115504	 (01) 00841506115504	Screw, Arthrodesis, 2.0 x 32mm, Ti HCSD-20032 (01)00841506115535	 (01) 00841506115535
Screw, Arthrodesis, 2.0 x 26mm, Ti HCSD-20026 (01)00841506115511	 (01) 00841506115511	Screw, Arthrodesis, 2.0 x 34mm, Ti HCSD-20034 (01)00841506115542	 (01) 00841506115542
Screw, Arthrodesis, 2.0 x 36mm, Ti HCSD-20036 (01)00841506117157	 (01) 00841506117157	Screw, Arthrodesis, 2.0 x 40mm, Ti HCSD-20040 (01)00841506108926	 (01) 00841506108926
Screw, Arthrodesis, 2.0 x 44mm, Ti HCSD-20044 (01)00841506117164	 (01) 00841506117164		
2.5mm Arthrodesis Screw			
Screw, Arthrodesis, 2.5 x 26mm, Ti HCSD-25026 (01)00841506115559	 (01) 00841506115559	Screw, Arthrodesis, 2.5 x 34mm, Ti HCSD-25034 (01)00841506115580	 (01) 00841506115580
Screw, Arthrodesis, 2.5 x 28mm, Ti HCSD-25028 (01)00841506115566	 (01) 00841506115566	Screw, Arthrodesis, 2.5 x 36mm, Ti HCSD-25036 (01)00841506115597	 (01) 00841506115597
Screw, Arthrodesis, 2.5 x 30mm, Ti HCSD-25030 (01)00841506108957	 (01) 00841506108957	Screw, Arthrodesis, 2.5 x 38mm, Ti HCSD-25038 (01)00841506115603	 (01) 00841506115603
Screw, Arthrodesis, 2.5 x 32mm, Ti HCSD-25032 (01)00841506115573	 (01) 00841506115573	Screw, Arthrodesis, 2.5 x 40mm, Ti HCSD-25040 (01)00841506108971	 (01) 00841506108971

REDUCT® Arthrodesis Screw System

Inventory Control Sheet

3.5mm Arthrodesis Screw			
Screw, Arthrodesis, 3.5 x 32mm, Ti HCSD-35032 (01)00841506115429	 (01)00841506115429	Screw, Arthrodesis, 3.5 x 40mm, Ti HCSD-35040 (01)00841506109039	 (01)00841506109039
Screw, Arthrodesis, 3.5 x 34mm, Ti HCSD-35034 (01)00841506115436	 (01)00841506115436	Screw, Arthrodesis, 3.5 x 42mm, Ti HCSD-35042 (01)00841506115467	 (01)00841506115467
Screw, Arthrodesis, 3.5 x 36mm, Ti HCSD-35036 (01)00841506115443	 (01)00841506115443	Screw, Arthrodesis, 3.5 x 44mm, Ti HCSD-35044 (01)00841506115474	 (01)00841506115474
Screw, Arthrodesis, 3.5 x 38mm, Ti HCSD-35038 (01)00841506115450	 (01)00841506115450	Screw, Arthrodesis, 3.5 x 46mm, Ti HCSD-35046 (01)00841506115481	 (01)00841506115481
Single Use (Disposable) Instruments			
K-Wire, 0.9mm x 127mm, Double Trocar KWIR-DT-09127 (01)00841506109060	 (01)00841506109060	K-Wire, HCS, 1.1mm x 127mm, Double Trocar KWIR-DT-11127 (01)00841506209098	 (01)00841506209098
Drill, 1.8mm Cannulated DRLL-DIP-18 (01)00841506112756	 (01)00841506112756	Drill, 2.1mm Cannulated DRLL-DIP-21 (01)00841506109237	 (01)00841506109237
Drill, 2.9mm Cannulated DRLL-DIP-29 (01)00841506109244	 (01)00841506109244	REDUCT Driver, 2.0 DRVR-HCS-0110 (01)00841506107288	 (01)00841506107288
REDUCT Driver, 2.5 DRVR-HCS-1015 (01)00841506109213	 (01)00841506109213	REDUCT Driver, 3.5 DRVR-HCS-1520 (01)00841506109220	 (01)00841506109220
Reusable Instruments			
Handle, Small QC, Fixed HNDL-SQC-FXD (01)00841506102078	 (01)00841506102078	Handle, AO QC, Fixed HNDL-AQC-FXD (01)00841506105406	 (01)00841506105406
HCS Wire Pusher HCS-WP (01)00841506105413	 (01)00841506105413	REDUCT Depth Gauge DGA-DIP (01)00841506109268	 (01)00841506109268

REDUCT® Arthrodesis Screw System

Inventory Control Sheet

Convex Reamer, Size 1 RMR-CX-01 (01)00841506109343 <div style="text-align: center;">  (01)00841506109343 </div>	Concave Reamer, Size 1 RMR-CV-01 (01)00841506109312 <div style="text-align: center;">  (01)00841506109312 </div>
Convex Reamer, Size 2 RMR-CX-02 (01)00841506109350 <div style="text-align: center;">  (01)00841506109350 </div>	Concave Reamer, Size 2 RMR-CV-02 (01)00841506109329 <div style="text-align: center;">  (01)00841506109329 </div>
Convex Reamer, Size 3 RMR-CX-03 (01)00841506109367 <div style="text-align: center;">  (01)00841506109367 </div>	Concave Reamer, Size 3 RMR-CV-03 (01)00841506109336 <div style="text-align: center;">  (01)00841506109336 </div>
Convex Reamer, Size 4 RMR-CX-04 (01)00841506114392 <div style="text-align: center;">  (01)00841506114392 </div>	Concave Reamer, Size 4 RMR-CV-04 (01)00841506114378 <div style="text-align: center;">  (01)00841506114378 </div>
Convex Reamer, Size 5 RMR-CX-05 (01)00841506114408 <div style="text-align: center;">  (01)00841506114408 </div>	Concave Reamer, Size 5 RMR-CV-05 (01)00841506114385 <div style="text-align: center;">  (01)00841506114385 </div>
Skin Hook, 2 Prong, 2mm SH2-020 (01)00841506114569 <div style="text-align: center;">  (01)00841506114569 </div>	